



CHPV-C 2023 Human Papillomavirus

INSTITUTION: Direccion de Redes Integradas de Salud Lima Este
Lima Lim PE 15007

ATTENTION: Yasminda Cajacuri Sanchez

CAP NUMBER: 9044993-01 **Kit#** 1

KIT INFORMATION:	Kit ID:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	37102184	12/11/2023	1/16/2024	2/26/2024

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
[26] = Educational challenge.

Reviewed By

Date



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 9044993-01

Kit# 1

Kit ID: 37102184

Institution: Direccion de Redes Integradas de Salud Lima Este

Kit Mailed: 12/11/2023

Attention: Yasmina Cajacuri Sanchez

Original Evaluation: 1/16/2024

City / State: Lima Lim PE 15007

**EVALUATION
ORIGINAL**

CHPV-C 2023 Human Papillomavirus

Test Method	Specimen	Your Result	Intended Response	Your Grade
HRHPV ROCHE COBAS 4800	CHPVM-11	NEGATIVE	NEGATIVE	Good
	CHPVM-12	POSITIVE	POSITIVE	Good
	CHPVM-13	NEGATIVE	NEGATIVE	Good
	CHPVM-14	POSITIVE	POSITIVE	Good
	CHPVM-15	NEGATIVE	NEGATIVE	Good
HR HPV Genotyping	CHPVM-11	HIGH RISK HPV NEGATIVE		See Note [26]
	CHPVM-12	HPV 18		See Note [26]
	CHPVM-13	HIGH RISK HPV NEGATIVE		See Note [26]
	CHPVM-14	HPV 16		See Note [26]
	CHPVM-15	HIGH RISK HPV NEGATIVE		See Note [26]

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 9044993-01

Kit# 1

Kit ID: 37102184

Institution: Direccion de Redes Integradas de Salud Lima Este

Kit Mailed: 12/11/2023

Attention: Yasminda Cajacuri Sanchez

Original Evaluation: 1/16/2024

City / State: Lima Lim PE 15007

**EVALUATION
ORIGINAL**

CHPV-C 2023 Human Papillomavirus

CAP

CAP #: 9044993

Subspecialty : Virology

Regulated Analyte	Proficiency Event 2023 1				Proficiency Event 2023 2				Proficiency Event 2023 3			
	Test Event	Score	%	Wgt	Test Event	Score	%	Wgt	Test Event	Score	%	Wgt
Viral Identification	CHPV-A	5/5	100	5	CHPV-B	5/5	100	5	CHPV-C	5/5	100	5
Performance By Mailing	Satisfactory 100				Satisfactory 100				Satisfactory 100			
Regulatory Status for Virology	Cumulative Regulatory Performance Interpretation: Successful											

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.